




## Calibration Report

Part Number :	<b>DCL-05-TCP</b>
Description :	<b>Single Axis Closed Loop Inclinometer</b>
Accuracy @ 20°C :	<b>±0.0025</b>
Serial Number :	<b>49832</b>
Date :	<b>19/11/2016</b>

### Calibration Test - Readings in Degrees

Cal point	Angle	Error
-5.0000	-4.9997	-0.0003
-4.5000	-4.4999	-0.0001
-4.0000	-3.9997	-0.0003
-3.5000	-3.5002	0.0002
-3.0000	-3.0000	0.0000
-2.5000	-2.4999	-0.0001
-2.0000	-1.9999	-0.0001
-1.5000	-1.4999	-0.0001
-1.0000	-0.9999	-0.0001
-0.5000	-0.4998	-0.0002
0.0000	0.0000	0.0000
0.5000	0.4999	0.0001
1.0000	1.0001	-0.0001
1.5000	1.5001	-0.0001
2.0000	2.0002	-0.0002
2.5000	2.5004	-0.0004
3.0000	3.0001	-0.0001
3.5000	3.5002	-0.0002
4.0000	4.0004	-0.0004
4.5000	4.5005	-0.0005
5.0000	5.0004	-0.0004

Test Voltage :	<b>24Vdc</b>
Measurement taken by :	<b>Jesper Sotikov</b>
Equipment Used :	<b>ITCS &amp; DCL App</b>
Signed :	



level developments

Level Developments Ltd  
Spencer Place  
97-99 Gloucester Road  
Croydon  
Surrey CR0 2DN  
United Kingdom

t. +44 (0)20 86841400  
f. +44 (0)20 86841422  
sales@leveldevelopments.com

[www.leveldevelopments.com](http://www.leveldevelopments.com)

### 49832 Error Over Measurement Range

