



Calibration Report

Part Number :	VS-10-C-1-0
Description :	Dual Axis Inclinometer Sensor ±10° Range
Accuracy @ 20°C :	±0.03°
Serial Number :	52352
Date :	11/11/2016

Calibration Test - Readings in Degrees

		Y Axis				
Angle		-10.00	-5.00	0.00	5.00	10.00
X Axis	-10.00	-10.003 -10.003		-0.003 -9.990		10.009 -10.000
	-5.00		-5.002 -4.995	-0.005 -4.991	4.996 -4.993	
	0.00	-10.000 0.002	-4.996 0.000	0.000 0.000	4.994 0.001	9.993 0.002
	5.00		-4.997 4.999	-0.012 4.998	4.989 5.012	
	10.00	-9.998 10.002		-0.001 9.997		10.006 10.018

Test Voltage :	24Vdc
Measurement taken by :	Jesper Sotikov
Signed :	