



Calibration Report

Part Number :	VS-15-C-2-0
Description :	Dual axis, ±15°, RS232, M9 connector with TC
Accuracy @ 20°C :	±0.03°
Serial Number :	135429
Date :	27/07/2021

Calibration Test - Readings in Degrees

		Y Axis				
Angle		-15.00	-10.00	0.00	10.00	15.00
X Axis	-15.00	-15.002 -15.004		0.001 -15.004		15.003 -15.002
	-10.00		-10.002 -10.002	0.000 -10.001	10.003 -10.000	
	0.00	-15.003 0.000	-10.002 0.000	0.000 0.000	10.003 0.000	15.006 -0.001
	10.00		-10.000 10.000	0.001 10.000	10.003 9.998	
	15.00	-15.001 14.999		0.000 14.999		15.004 14.997

Test Voltage :	24Vdc
Equipment Used :	Tektronix DMM4040 (SN0922) Tektronix DMM4040 (SN0923)
Measurement taken by :	
Signed :	